



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/701,455	11/06/2003	Eric B. Stenzel	12013/48301	8547

23838 7590 04/19/2007
KENYON & KENYON LLP
1500 K STREET N.W.
SUITE 700
WASHINGTON, DC 20005

EXAMINER

GHERBI, SUZETTE JAIME J

ART UNIT	PAPER NUMBER
----------	--------------

3738

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
2 MONTHS	04/19/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

**UNITED STATES DEPARTMENT OF COMMERCE****U.S. Patent and Trademark Office**

Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450

APPLICATION NO./ CONTROL NO.	FILING DATE	FIRST NAMED INVENTOR / PATENT IN REEXAMINATION	ATTORNEY DOCKET NO.
---------------------------------	-------------	---	---------------------

EXAMINER

ART UNIT	PAPER
----------	-------

20061204

DATE MAILED:

Please find below and/or attached an Office communication concerning this application or proceeding.

Commissioner for Patents

The Supplemental Examiners Answer is to reflect the patent numbers used in the Evidence Section. No other changes have been made.



UNITED STATES PATENT AND TRADEMARK OFFICE

Commissioner for Patents
United States Patent and Trademark Office
P.O. Box 1450
Alexandria, VA 22313-1450
www.uspto.gov

**BEFORE THE BOARD OF PATENT APPEALS
AND INTERFERENCES**

Application Number: 10/701,455
Filing Date: November 06, 2003
Appellant(s): STENZEL, ERIC B.

Douglas E. Ringel
For Appellant

EXAMINER'S ANSWER

This is in response to the appeal brief filed September 13, 2006 appealing from the Office action mailed September 23, 2005.

(1) Real Party in Interest

A statement identifying by name the real party in interest is contained in the brief.

(2) Related Appeals and Interferences

The examiner is not aware of any related appeals, interferences, or judicial proceedings, which will directly affect or be directly affected by or have a bearing on the Board's decision in the pending appeal.

(3) Status of Claims

The statement of the status of claims contained in the brief is correct.

(4) Status of Amendments After Final

No amendment after final has been filed.

(5) Summary of Claimed Subject Matter

The summary of claimed subject matter contained in the brief is correct.

(6) Grounds of Rejection to be Reviewed on Appeal

Art Unit: 3738

The appellant's statement of the grounds of rejection to be reviewed on appeal is correct.

(7) Claims Appendix

The copy of the appealed claims contained in the Appendix to the brief is correct.

(8) Evidence Relied Upon

6849089	Stoll	2-2005
6339130	Bennett et al.	1-2002

(9) Grounds of Rejection

The following ground(s) of rejection are applicable to the appealed claims:

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1, 4, 9-10, 12-16, 24-27 are rejected under 35 U.S.C. 103(a) as being obvious over Stoll 6,849,089. Stoll discloses the invention as claimed noting figure 7 comprising: A medical structure (2); a set of first coated pellets (the pellet is the internal

Art Unit: 3738

structure 4" and the coating is the thin surrounding) each of the first coated pellets containing at least one first therapeutic composition, the set of first coated pellets deposited on the structure at a first site for controlled delivery of at least one first therapeutic composition; and a set of second coated pellets, (the second set of pellets can be interpreted as the internal pellet 4" with the thick outer diameter coating adjacent pellets 4" with the thin coating as seen in figure 7) each of the second coated pellets containing at least one second therapeutic composition (the second therapeutic composition is contained in the micro-sphere in a separate second compartment; the set of second coated pellets deposited on the structure at a second site for controlled delivery of the at least one second therapeutic composition to a desired location within the body; wherein each of said first coated pellets is covered with a first coating (the thin coating) and each of said second coated pellets is covered with a second coating (the thick coating); wherein the first coating is thinner than the second coating and has a faster in vivo decomposition rate relative to the second coating to release the first therapeutic composition from the first site faster than the second therapeutic composition from the second site (see col. 13, lines 18-29 and col. 4, lines 1-27); wherein each of the first coated pellets contains a substance in addition to the first therapeutic (see col. 12, lines that states that a "matrix" of a polymer and particles of cell proliferation-inhibiting substance may be used therefore the additional substance is the matrix blend). It is obvious to one having ordinary skill in the art that a matrix comprises more than one substance.

Claims 5, 11, 17, 19-23 are rejected under 35 U.S.C. 103(a) as being unpatentable over Stoll in view of Bennett et al. 6,339,130. Stoll has been disclosed above however Stoll's adhesive is not described. Bennett et al. teaches that adhesives for use with prosthetic devices are well known and polymers are known adhesive with curing properties (see abstract and col. 3, lines 12-35). It would have been obvious to one having ordinary skill in the art at the time the invention was made to make the adhesive of Stoll with the claimed properties such as polymers because Bennett et al. teaches that they can be used with a variety of surgical devices. Stoll also does not specify that a plurality of sublayers on the micro-spheres. It would have been obvious to one having ordinary skill in the art at the time the invention was made to add a plurality of sublayers because Stoll does note that a plurality of layers are contemplated (see figure 5) in order to vary the degradation and release rate of the therapeutic compositions.

(10) Response to Argument

Applicant's arguments, filed 9/13/04, with respect to the 112 first paragraph rejection have been fully considered and are persuasive. The rejection of claims 1, 4-5, 9-17, 19-21, 23-27 under 112 first paragraph has been withdrawn.

Claims 1,4-5,9-17, 19-21, 23-27 remain rejected under 35 USC §103. Applicant contends that Stoll does not disclose or suggest pellets having different coating

Art Unit: 3738


thickness. The examiner has pointed out figure 7 where clearly illustrations that the walls (5'') have varying wall thicknesses. Applicant contends that that Stoll does not disclose an additional "substance". This terminology is broad and is being interpreted as any additional material therefore the "matrix" as disclosed by Stoll does meet the claim limitation. Therefor the substance in addition to the first therapeutic composition if the make-up of the bio-degradable polymer.

(11) Related Proceeding(s) Appendix

No decision rendered by a court or the Board is identified by the examiner in the Related Appeals and Interferences section of this examiner's answer.

For the above reasons, it is believed that the rejections should be sustained.

Respectfully submitted,



SUZETTE GHERBI
PRIMARY EXAMINER
TECHNOLOGY CENTER 3700

Suzette J-J Gherbi 3-21-07

Conferees:



CORRINE McDERMOTT
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700



ANGELA D. SYKES
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700